

Declaration of Conformity

GKE Steri-Record® Incubators

Art. No.: 610-119/ -120/ -121/ -122
incl. aluminum block for self-contained biological indicators (SCBI)

Art. No.: 610-109/ -110/ -111/ -112
to be used with one of the following aluminum blocks:

for SCBI	(art. no.: 610-113)
for Stearo ampoules	(art. no.: 610-114)
for growth media	(art. no.: 610-115)

The incubator consists of an electrical heating device and an aluminum block which is electrically heated with a low voltage power supply to incubate biological indicators.

GKE hereby declares that the incubator is in compliance with the requirements of

Low Voltage Directive 2014/35/EU

EN 61010-1
EN 61010-2

Electromagnetic Compatibility (EMC) 2014/30/EU

EN 61000-6-3
EN 61000-6-1
EN 61326-1

RoHS Directive 2011/65/EU

We hereby declare under our sole responsibility that the GKE Steri-Record® incubator as defined above has been tested according to the above standard and thus meets all requirements to allow CE-marking.

Waldems, 2021-08-02

A blue ink signature of Ulrich Kaiser, consisting of a stylized 'U' and 'K' followed by a horizontal line.

Dipl.-Ing. Dr. Ulrich Kaiser
R & D-Manager

752-004-EN V09